

## Clarifications and Corrigendum Tender Notice 1 for Chemistry Lab Equipments

SB/QEL/PUR/TENDER/INST & ACCYS-1/2017/Corrigendum

2-11-2017

### Clarification:

With reference to the Tender Notice SB/QEL/PUR/TENDER/INST & ACCYS-1/2017 date 9-10-2017, in response to the queries received, it is hereby clarified that the following instruments in the tender require 5 year performance guarantee, as stated in the terms and conditions of the tender (page 5, point 20):

Annexure	Instrument	Performance Bank Guarantee	
		No. of Years	Amount
Annexure 1	GC MS/MS System	5	20% of order value
Annexure 2	Recirculating Chiller	5	20% of order value
Annexure 3	Rotary Evaporator	5	20% of order value
Annexure 5	Electrochemical Derivatization System for analysis of Aflatoxins	5	20% of order value
Annexure 7	LC MS/MS System	5	20% of order value
Annexure 8	Microwave Digester	5	20% of order value
Annexure 9	Water Purification System	5	20% of order value
Annexure 14	GC with FID	5	20% of order value
Annexure 15	GC MS system	5	20% of order value
Annexure 18	Concentration workstation	5	20% of order value
Annexure 19	Portable nitrogen generator for laboratory benchtop applications	5	20% of order value
Annexure 20	Refrigerated centrifuge	5	20% of order value

### Corrigendum:

1. It is hereby informed that item No. 7 , LC MS/MS system, **is withdrawn from the present tender**. This tender will be called for again at a later date.
2. Annexure 5, Electrochemical Derivatization System for Analysis of Aflatoxins:
  1. The title may be read as: *Electrochemical Derivatization System (or equivalent system) for Analysis of Aflatoxins*
  2. Section II, Required Specification, point no. 1 may be read as: *Post column electrochemical derivatization unit (or equivalent system) for derivatization of Aflatoxins B1 and G1 for analysis by HPLC*
  3. Point No. 7, warranty clause, may be read as: ***The system should be covered under comprehensive warranty for five years, inclusive of spares including the reactor coil or membrane in the system.***
3. Annexure 22, Custom QuEChERS Kits, Section II
  1. Required Specification no. 1 may be read as: *Sachets or dSPE tubes of the two sets of QuEChERS chemicals, with compositions as detailed in Table 2 below, is to be quoted.*
  2. Required Specification no. 5 should be read as: *The weighing should be accurate up to 5% of total weight and this should be indicated in the certificate.*

3. Required Specification no. 6, Labelling requirements, after the present text, the following is to be added: ***In case of custom labelling difficulties on small dSPE tubes, the tubes should be distinguishable by code numbers or colour coding.***
  4. Required Specification no. 11, in addition to the present text, the following to be added: ***If samples for the custom compositions cannot be provided along with the tender, then a statement of undertaking to immediately replace any product found defective should accompany the tender.***
  4. Annexure 9: Water Purification System, Section II, Required Specification no. 11 should be read as: *Stand alone point of delivery unit, **capable of delivering a flow of minimum 60L/hr (or higher)**, with...*
-